From: Misty Haynes
Sent: Wed, 12 Sep 2012 15:01:29 +0000
To: Annette Robinson <arobinson@neccrx.com>
Subject: Sterility Result

Hi Annette,

The positive sterility result from lot # 08272012@87 (ARL # 185744) is *not* being invalidated. The environmental monitoring from that day does not show a contamination that matches the sample contamination. I have issued a Not Sterile report for this sample.

If you would like us to Gram Stain the contamination, please let me know. Bacterial identification is also available, upon your request. If you want either of these services we will be happy to take care of them for you, otherwise this OOS will be closed.

Thank you,

Misty Haynes

Microbiologist

ARL